

# Certificate of Participation

for the EURADOS Intercomparison 2022 for whole body dosimeters (IC2022ph)

<b>Certificate Number:</b>	EURADOS-IC2022ph-S050 for system S050/2022
<b>Number of pages:</b>	4
<b>Date of Issue:</b>	July 07 <sup>th</sup> , 2023
<b>Participating Institute:</b>	TECNORAD s.u.r.l., Italy
<b>Dosimetry System:</b>	Green Film
<b>Reporting number:</b>	111 (this anonymous number will be used in further publications)
<b>Intercomparison procedure:</b>	<p>The EURADOS Intercomparison 2022 for whole body dosimeters was managed and coordinated on behalf of EURADOS by the WG2 Intercomparison Organization Group (OG). The OG established the irradiation plan and announced the intercomparison, including the range limits of the doses and radiation qualities, in May 2022.</p> <p>Participants were asked to indicate details of the dosimeter reference point on the online application form. After completing application procedures the participants sent their dosimeters, according to the instructions, to the OG coordinator (July 2022). The coordinator checked the correct labelling of the dosimeters and transferred all dosimeters, along with the instructions, to the irradiation laboratory. The laboratory irradiated the dosimeters according to the irradiation plan and sent all the dosimeters back to the coordinator (November 2022).</p> <p>The coordinator then returned the dosimeters to the participants for assessment and indicated which dosimeters were not irradiated. The participants were instructed to follow normal routine procedures as far as possible. The participants then sent the results of the dosimeter readings to the coordinator (January 2023). After receipt of the participants' results, the coordinator sent the irradiation data to the participants.</p>
<b>Number of participants:</b>	96 institutes participated in IC2022ph with a total of 116 systems.
<b>Coordinator:</b>	Christian Gärtner (Seibersdorf Labor GmbH, A-2444 Seibersdorf)
<b>Intercomparison results:</b>	See the tables on pages 2 to 4 of this certificate.
<b>Irradiation data:</b>	See the attached certificate of the irradiation laboratory: Number Dos/2896-050/2022
<b>Participant results:</b>	See the attached signed dose report provided by the participant.

On behalf of the intercomparison  
Organization Group:



Christian Gärtner  
Coordinator

On behalf of EURADOS:



Filip Vanhavere  
Chairperson

## Whole body dosemeter intercomparison IC2022ph

### Result of the Intercomparison (Dosimetry System S050/2022)

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality	Quantity	Participant's Value	Reference Value	Ratio
S050/2022-01	S050/2022-01	S-Cs, 0°	$H_p(10)$	2.77 mSv	2.70 mSv	1.02
			$H_p(0.07)$	2.77 mSv	2.70 mSv	1.02
S050/2022-07	S050/2022-07	S-Cs, 0°	$H_p(10)$	2.72 mSv	2.70 mSv	1.01
			$H_p(0.07)$	2.72 mSv	2.70 mSv	1.01
S050/2022-12	S050/2022-12	S-Cs, 0°	$H_p(10)$	2.76 mSv	2.70 mSv	1.02
			$H_p(0.07)$	2.76 mSv	2.70 mSv	1.02
S050/2022-27	S050/2022-27	S-Cs, 0°	$H_p(10)$	2.85 mSv	2.70 mSv	1.06
			$H_p(0.07)$	2.85 mSv	2.70 mSv	1.06
S050/2022-06	S050/2022-06	S-Cs, 0°	$H_p(10)$	9.16 mSv	8.80 mSv	1.04
			$H_p(0.07)$	9.16 mSv	8.80 mSv	1.04
S050/2022-17	S050/2022-17	S-Cs, 0°	$H_p(10)$	9.34 mSv	8.80 mSv	1.06
			$H_p(0.07)$	9.34 mSv	8.80 mSv	1.06
S050/2022-04	S050/2022-04	S-Co, 0°	$H_p(10)$	9.58 mSv	8.50 mSv	1.13
			$H_p(0.07)$	9.75 mSv	8.65 mSv	1.13
S050/2022-15	S050/2022-15	S-Co, 0°	$H_p(10)$	9.48 mSv	8.50 mSv	1.12
			$H_p(0.07)$	9.65 mSv	8.65 mSv	1.12
S050/2022-11	S050/2022-11	S-Co, 0°	$H_p(10)$	127.10 mSv	122.00 mSv	1.04
			$H_p(0.07)$	129.40 mSv	124.10 mSv	1.04
S050/2022-13	S050/2022-13	S-Co, 0°	$H_p(10)$	126.70 mSv	122.00 mSv	1.04
			$H_p(0.07)$	128.90 mSv	124.10 mSv	1.04

### Whole body dosimeter intercomparison IC2022ph

#### Result of the Intercomparison (Dosimetry System S050/2022), continued

EURADOS Dosemeter ID	Participant's Dosemeter ID	Radiation Quality	Quantity	Participant's Value	Reference Value	Ratio
S050/2022-16	S050/2022-16	S-Co, 0°	$H_p(10)$	258.10 mSv	281.00 mSv	0.92
			$H_p(0.07)$	262.60 mSv	285.90 mSv	0.92
S050/2022-23	S050/2022-23	S-Co, 0°	$H_p(10)$	261.70 mSv	281.00 mSv	0.93
			$H_p(0.07)$	266.20 mSv	285.90 mSv	0.93
S050/2022-02	S050/2022-02	N-40, 0°	$H_p(10)$	5.74 mSv	5.57 mSv	1.03
			$H_p(0.07)$	6.08 mSv	5.89 mSv	1.03
S050/2022-18	S050/2022-18	N-40, 0°	$H_p(10)$	5.60 mSv	5.57 mSv	1.01
			$H_p(0.07)$	5.92 mSv	5.89 mSv	1.01
S050/2022-05	S050/2022-05	N-40 + S-Cs mixed, 0°	$H_p(10)$	3.87 mSv	3.40 mSv	1.14
			$H_p(0.07)$	3.98 mSv	3.50 mSv	1.14
S050/2022-08	S050/2022-08	N-40 + S-Cs mixed, 0°	$H_p(10)$	3.84 mSv	3.40 mSv	1.13
			$H_p(0.07)$	3.95 mSv	3.50 mSv	1.13
S050/2022-03	S050/2022-03	W-80, 0°	$H_p(10)$	6.56 mSv	6.30 mSv	1.04
			$H_p(0.07)$	6.08 mSv	5.84 mSv	1.04
S050/2022-28	S050/2022-28	W-80, 0°	$H_p(10)$	6.65 mSv	6.30 mSv	1.06
			$H_p(0.07)$	6.16 mSv	5.84 mSv	1.05
S050/2022-09	S050/2022-09	W-80, 60°	$H_p(10)$	5.20 mSv	5.90 mSv	0.88
			$H_p(0.07)$	4.89 mSv	6.37 mSv	0.77
S050/2022-10	S050/2022-10	W-80, 60°	$H_p(10)$	5.22 mSv	5.90 mSv	0.88
			$H_p(0.07)$	4.90 mSv	6.37 mSv	0.77

**Whole body dosimeter intercomparison IC2022ph**

**Result of the Intercomparison (Dosimetry System S050/2022), continued**

<b>EURADOS Dosemeter ID</b>	<b>Participant's Dosemeter ID</b>	<b>Radiation Quality</b>
S050/2022-14	S050/2022-14	not irradiated
S050/2022-19	S050/2022-19	not irradiated
S050/2022-20	S050/2022-20	not irradiated
S050/2022-21	S050/2022-21	not irradiated
S050/2022-22	S050/2022-22	not irradiated
S050/2022-24	S050/2022-24	not irradiated
S050/2022-25	S050/2022-25	not irradiated
S050/2022-26	S050/2022-26	not irradiated
S050/2022-29	S050/2022-29	not irradiated
S050/2022-30	S050/2022-30	not irradiated

Radiation Qualities and average photon energy (according to ISO 4037-1 and IEC 61267):

- Nuclide Radiation:
  - S-Cs: 662 keV
  - S-Co: 1250 keV
- X-Rays:
  - N-40: 33 keV
  - W-80: 56.5 keV



Calibrations  
Cert. No 116(s)

HELLENIC REPUBLIC  
MINISTRY OF DEVELOPMENT AND INVESTMENTS  
GENERAL SECRETARIAT FOR RESEARCH & INNOVATION

Ag. Paraskevi, 19/05/2023  
Our Ref: B/428/11274



IONIZING RADIATION CALIBRATION LABORATORY  
Affiliated to the Hellenic Metrology Institute

IRRADIATION CERTIFICATE No: Dos/2896- 050/2022  
Number of Pages: 2  
Date of Issue: 24/05/2023

The following personnel dosimeters from:

EURADOS INTERCOMPARISON PROGRAM  
SYSTEM No:S050/2022

have been irradiated at the *Ionizing Radiation Calibration Laboratory of Greek Atomic Energy Commission*:

Personal Dosimeters (PD):	Whole body
Dosimeter Identification:	-
Detection Principle:	-
Irradiation Period:	See below

The Kair reference values have been obtained using the secondary standards ionization chambers PTW W-32002-LS01 (S/N: 69) or FC65-G (S/N: 634) and the electrometer PTW UNIDOS 10002 (S/N 20314). The LS01 chamber was calibrated in PTB during 06 & 07-03-2019 (PTB, Cal. Cert. No PTB-6.3-4094018). The FC65-G chamber was calibrated at BIPM on 24-04-2019 (BIPM, Cal. Cert. No 25). The irradiation conditions are in accordance to ISO 4037/1-2-3-4 and IEC 62387.

**Irradiation conditions**

Phantom:	ISO water phantom, (30x30x15) cm <sup>3</sup>
Source to PD Distance:	100-300 cm, depending on required Kair rate
Kair Rate:	S-Cs: 454.3 µGy/min (at 100 cm) W-80: 2.63 mGy/min (at 200 cm) N-40: 203.7 µGy/min (at 200 cm) S-Co: 145.5 mGy/min (at 100 cm) S-Co: 18.6 mGy/min (at 300 cm) S-Co: 0.90 mGy/min (at 300 cm with lead block)
Field Size:	S-Cs: Circular with diameter of 55.6 cm (at 200 cm) x-rays: Circular with diameter 26.8 cm (at 200 cm) S-Co: Rectangular (30x30) cm <sup>2</sup> (at 300 cm)
Build up PMMA:	(0.3 x 30x30) cm <sup>3</sup>
Reference point of PD:	Frontal surface of slab phantom

**Environmental conditions during irradiations:**

Temperature	Pressure	Relative Humidity
19.0-21.0 °C	981.0-985.0 hPa	10 %

T.Θ. 60092, Agia Paraskevi 153 10 Attiki, Tel. : +30 210 650 6765  
e-mail : argiro.boziari@eeae.gr, info@eeae.gr

**Irradiation Data**

# Dosemeter	Date	Quality	H <sub>p</sub> (10) mSv	U % <sup>(1)</sup>	H <sub>p</sub> (0.07) mSv	U % <sup>(1)</sup>
S050-12	09/09/2022	S-Cs	2.70	4.9	2.70	4.9
S050-27	09/09/2022	S-Cs	2.70	4.9	2.70	4.9
S050-07	09/09/2022	S-Cs	2.70	4.9	2.70	4.9
S050-01	09/09/2022	S-Cs	2.70	4.9	2.70	4.9
S050-06	30/08/2022	S-Cs	8.80	4.9	8.80	4.9
S050-17	30/08/2022	S-Cs	8.80	4.9	8.80	4.9
S050-04	16/11/2022	S-Co	8.50	4.9	8.65	4.9
S050-15	16/11/2022	S-Co	8.50	4.9	8.65	4.9
S050-11	22/09/2022	S-Co	122	4.9	124	4.9
S050-13	22/09/2022	S-Co	122	4.9	124	4.9
S050-16	30/08/2022	S-Co	281	4.9	286	4.9
S050-23	30/08/2022	S-Co	281	4.9	286	4.9
S050-18	29/09/2022	N-40-(0°)	5.57	5.1	5.89	5.1
S050-02	29/09/2022	N-40-(0°)	5.57	5.1	5.89	5.1
S050-28	04/11/2022	W-80-(0°)	6.30	5.1	5.84	5.1
S050-03	04/11/2022	W-80-(0°)	6.30	5.1	5.84	5.1
S050-09	10/11/2022	W-80-(60°)	5.90	5.3	6.37	5.3
S050-10	10/11/2022	W-80-(60°)	5.90	5.3	6.37	5.3

# Dosemeter	Date	Quality	H <sub>p</sub> (10) mSv	U % <sup>(1)</sup>	H <sub>p</sub> (0.07) mSv	U % <sup>(1)</sup>
S050-05	12/09/2022	Cs-137	1.60	4.9 %	1.60	4.9 %
	14/10/2022	N-40	1.80	5.1 %	1.90	5.1 %
	<b>Total</b>		<b>3.40</b>	<b>7.1 %</b>	<b>3.50</b>	<b>7.1 %</b>
S050-08	12/09/2022	Cs-137	1.60	4.9 %	1.60	4.9 %
	14/10/2022	N-40	1.80	5.1 %	1.90	5.1 %
	<b>Total</b>		<b>3.40</b>	<b>7.1 %</b>	<b>3.50</b>	<b>7.1 %</b>

<sup>1</sup>U= uncertainty 95% confidence level (k=2)

**Not Irradiated Dosimeters :**

S050-14; S050-19; S050-20; S050-21; S050-22; S050-24; S050-25; S050-26; S050-29; S050-30;

Irradiations performed by:

Boziari A., Medical Physicist  
Stamatopoulou E., Technician  
Askounis P., Physicist



**Argiro Boziari**  
**Head of Dosimetry and Calibration Unit**



This certificate is issued in accordance with the requirements of ISO 17025. It provides traceability of measurements to recognized national standards laboratories. The HIRCL/GAEC is a member of the IAEA/WHO Secondary Standard Dosimetry Laboratory Network. This certificate may not be reproduced other than in full, except with the prior written approval of the HIRCL/GAEC

T.Θ. 60092, Agia Paraskevi 153 10 Attiki, Tel. : +30 210 650 6765

e-mail : argiro.boziari@eeae.gr, info@eeae.gr

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

**P051/2022, TECNORAD s.u.r.l., 37135 VERONA (Italy)**

**S050/2022, Green Film**

*Dose Values*

<b>Dosemeter 'S050/2022 - 01':</b>	<b>Hp(10) dose = 2.765 mSv</b>
S050/2022-01	<b>Hp(0.07) dose = 2.765 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 02':</b>	<b>Hp(10) dose = 5.744 mSv</b>
S050/2022-02	<b>Hp(0.07) dose = 6.076 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 03':</b>	<b>Hp(10) dose = 6.557 mSv</b>
S050/2022-03	<b>Hp(0.07) dose = 6.075 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 04':</b>	<b>Hp(10) dose = 9.584 mSv</b>
S050/2022-04	<b>Hp(0.07) dose = 9.751 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 05':</b>	<b>Hp(10) dose = 3.868 mSv</b>
S050/2022-05	<b>Hp(0.07) dose = 3.976 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 06':</b>	<b>Hp(10) dose = 9.162 mSv</b>
S050/2022-06	<b>Hp(0.07) dose = 9.162 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 07':</b>	<b>Hp(10) dose = 2.719 mSv</b>
S050/2022-07	<b>Hp(0.07) dose = 2.719 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 08':</b>	<b>Hp(10) dose = 3.843 mSv</b>
S050/2022-08	<b>Hp(0.07) dose = 3.950 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 09':</b>	<b>Hp(10) dose = 5.201 mSv</b>
S050/2022-09	<b>Hp(0.07) dose = 4.887 mSv</b>
irradiated	<b>Remark =</b>
<b>Dosemeter 'S050/2022 - 10':</b>	<b>Hp(10) dose = 5.216 mSv</b>
S050/2022-10	<b>Hp(0.07) dose = 4.902 mSv</b>
irradiated	<b>Remark =</b>

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

**P051/2022, TECNORAD s.u.r.l., 37135 VERONA (Italy)**

**S050/2022, Green Film**

*Dose Values, continued*

<b>Dosimeter 'S050/2022 - 11':</b>	Hp(10) dose = 127.100 mSv
S050/2022-11	Hp(0.07) dose = 129.400 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 12':</b>	Hp(10) dose = 2.756 mSv
S050/2022-12	Hp(0.07) dose = 2.756 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 13':</b>	Hp(10) dose = 126.700 mSv
S050/2022-13	Hp(0.07) dose = 128.900 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 14':</b>	Hp(10) dose = 0.000 mSv
S050/2022-14	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.218 mSv
<b>Dosimeter 'S050/2022 - 15':</b>	Hp(10) dose = 9.480 mSv
S050/2022-15	Hp(0.07) dose = 9.645 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 16':</b>	Hp(10) dose = 258.100 mSv
S050/2022-16	Hp(0.07) dose = 262.600 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 17':</b>	Hp(10) dose = 9.341 mSv
S050/2022-17	Hp(0.07) dose = 9.341 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 18':</b>	Hp(10) dose = 5.598 mSv
S050/2022-18	Hp(0.07) dose = 5.922 mSv
irradiated	Remark =
<b>Dosimeter 'S050/2022 - 19':</b>	Hp(10) dose = 0.000 mSv
S050/2022-19	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.221 mSv
<b>Dosimeter 'S050/2022 - 20':</b>	Hp(10) dose = 0.000 mSv
S050/2022-20	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.217 mSv

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

<i>General Data</i>	
<b>P051/2022, TECNORAD s.u.r.l., 37135 VERONA (Italy)</b>	
<i>Device</i>	
<b>S050/2022, Green Film</b>	

<i>Dose Values, continued</i>	
<b>Dosemeter 'S050/2022 - 21':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-21	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.228 mSv
<b>Dosemeter 'S050/2022 - 22':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-22	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.187 mSv
<b>Dosemeter 'S050/2022 - 23':</b>	<b>Hp(10) dose = 261.700 mSv</b>
S050/2022-23	Hp(0.07) dose = 266.200 mSv
irradiated	Remark =
<b>Dosemeter 'S050/2022 - 24':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-24	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.190 mSv
<b>Dosemeter 'S050/2022 - 25':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-25	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.189 mSv
<b>Dosemeter 'S050/2022 - 26':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-26	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.203 mSv
<b>Dosemeter 'S050/2022 - 27':</b>	<b>Hp(10) dose = 2.849 mSv</b>
S050/2022-27	Hp(0.07) dose = 2.849 mSv
irradiated	Remark =
<b>Dosemeter 'S050/2022 - 28':</b>	<b>Hp(10) dose = 6.648 mSv</b>
S050/2022-28	Hp(0.07) dose = 6.160 mSv
irradiated	Remark =
<b>Dosemeter 'S050/2022 - 29':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-29	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.224 mSv
<b>Dosemeter 'S050/2022 - 30':</b>	<b>Hp(10) dose = 0.000 mSv</b>
S050/2022-30	Hp(0.07) dose = 0.000 mSv
not irradiated	Remark = Background; 0.201 mSv

## Whole body dosimeter intercomparison IC2022ph Dose Values Form

*General Data*

P051/2022, TECNORAD s.u.r.l., 37135 VERONA (Italy)

S050/2022, Green Film

*Dose Values, continued*

Transit dose: Hp(10) dose = 0.208 mSv

Remark =

  
Signature



**TECNORAD**

PERSONAL DOSIMETRY SERVICE

TECNORAD s.u.r.l.

Via Schiaparelli, 5 - 37135 Verona (Italy)

Tel. +39 045 8201066 - info@tecnorad.it

C.F./P.I. e R.I.: 00645130238

DANIELE BONAMINI  
Name

21/01/2023  
Date